



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/047,482	10/23/2001	Kevin Richardson	1001.1441101	4029

28075 7590 12/22/2003

CROMPTON, SEAGER & TUFTE, LLC  
1221 NICOLLET AVENUE  
SUITE 800  
MINNEAPOLIS, MN 55403-2420

EXAMINER

KONTOS, LINA R

ART UNIT PAPER NUMBER

3763

DATE MAILED: 12/22/2003

Please find below and/or attached an Office communication concerning this application or proceeding.

# Office Action Summary

Application No.

10/047,482

Applicant(s)

RICHARDSON, KEVIN

Examiner

Lina Kontos

Art Unit

3763

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

## Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133).
- Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

## Status

- 1) ☒ Responsive to communication(s) filed on 22 September 2003.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

## Disposition of Claims

- 4) ☒ Claim(s) 1-28 is/are pending in the application.
- 4a) Of the above claim(s) 3,4,9-14,19-21,23 and 25-28 is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1,2,5-8,15-18 and 24 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

## Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

## Priority under 35 U.S.C. §§ 119 and 120

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).  
a) ☐ All b) ☐ Some \* c) ☐ None of:  
1. ☐ Certified copies of the priority documents have been received.  
2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.  
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).  
\* See the attached detailed Office action for a list of the certified copies not received.
- 13) ☐ Acknowledgment is made of a claim for domestic priority under 35 U.S.C. § 119(e) (to a provisional application) since a specific reference was included in the first sentence of the specification or in an Application Data Sheet. 37 CFR 1.78.  
a) ☐ The translation of the foreign language provisional application has been received.
- 14) ☐ Acknowledgment is made of a claim for domestic priority under 35 U.S.C. §§ 120 and/or 121 since a specific reference was included in the first sentence of the specification or in an Application Data Sheet. 37 CFR 1.78.

## Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892) 4) ☐ Interview Summary (PTO-413) Paper No(s). \_\_\_\_\_
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948) 5) ☐ Notice of Informal Patent Application (PTO-152)
- 3) ☐ Information Disclosure Statement(s) (PTO-1449) Paper No(s) \_\_\_\_\_ 6) ☐ Other: \_\_\_\_\_

Art Unit: 3763

***Information Disclosure Statement***

***1.***

While applicant has indicated that an IDS was submitted on January 25, 2002 and that the office received it, no IDS has been included in the file. Examiner has considered the US patents listed as references, but is unable to obtain the other art. Examiner requests resubmission of the Information Disclosure Statement with a copy of each reference

***Claim Rejections - 35 USC § 102***

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

***2.***

Claims 1,2,5-8,15,16 are rejected under 35 U.S.C. 102(b) as being anticipated by Yoon et al.

Yoon et al. teaches an expandable multifunction instrument comprising an outer tubular member (14) disposed concentrically around an inner tubular member (12). Attached to the inner tubular member is an expandable element (16) having a plurality of struts (40,44) is a preformed triangular shape. The expandable member is retained in its non-expanded position by the outer tube, and when the outer tube is removed from surrounding the expandable element, it springs to its expanded position (column 10, lines 41-46). The struts of the expandable element may be made from any suitable metal (column 10, line 15) and the material disposed over the

Art Unit: 3763

struts may be manufactured from many materials, including polymeric webbing or a semi-rigid polymer (column 5, lines 31-35).

### ***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

3.

Claims 17,18,22,24 are rejected under 35 U.S.C. 103(a) as being unpatentable over Yoon et al. in view of Karpiel et al.

Yoon et al., as described above, teaches an expandable multifunction instrument comprising an outer tubular member (14) disposed concentrically around an inner tubular member (12) having an expandable element (16) with a plurality of struts (40,44).

Karpiel teaches an ERCP catheter system having a tapered tip (16, column 5, line 23).

It would have been obvious to one skilled in the art at the time of the invention to have a tapered tip on the profile of the device in order to have a less obtrusive profile.

### ***Response to Arguments***

Applicant's arguments filed September 22, 2003 have been fully considered but they are not persuasive.

Art Unit: 3763

4.

In response to applicant's arguments, the recitation "biliary catheter for endoscopic retrograde cholangiopancreatography" has not been given patentable weight because the recitation occurs in the preamble. A preamble is generally not accorded any patentable weight where it merely recites the purpose of a process or the intended use of a structure, and where the body of the claim does not depend on the preamble for completeness but, instead, the process steps or structural limitations are able to stand alone. See *In re Hirao*, 535 F.2d 67, 190 USPQ 15 (CCPA 1976) and *Kropa v. Robie*, 187 F.2d 150, 152, 88 USPQ 478, 481 (CCPA 1951).

5.

In response to applicant's argument that the device of Yoon et al. creates a space where there previously had not been one, a recitation of the intended use of the claimed invention must result in a structural difference between the claimed invention and the prior art in order to patentably distinguish the claimed invention from the prior art. If the prior art structure is capable of performing the intended use, then it meets the claim. In a claim drawn to a process of making, the intended use must result in a manipulative difference as compared to the prior art. See *In re Casey*, 152 USPQ 235 (CCPA 1967) and *In re Otto*, 136 USPQ 458, 459 (CCPA 1963).

**Conclusion**

6.

Art Unit: 3763

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure.

US Patent 5,275,610 teaches a device for enlarging or maintaining a cavity having a radially expandable chute located at the distal end of a cannula.

**THIS ACTION IS MADE FINAL.** Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than **SIX MONTHS** from the mailing date of this final action.

Art Unit: 3763

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Lina Kontos whose telephone number is (703) 306-4207. The examiner can normally be reached on M-F 8:30-5:30.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Brian Casler can be reached on (703) 308-3552. The fax phone number for the organization where this application or proceeding is assigned is (703) 872-9302.

Any inquiry of a general nature or relating to the status of this application or proceeding should be directed to the receptionist whose telephone number is (703) 308-1148.

LRK

*LRK*

12-14-03



MICHAEL J. HAYES  
PRIMARY EXAMINER